




Performance of New Sampling Calorimeter in the KOTO Experiment

김준이, 이종원, 안정근(고려대), 김은주(전북대), 임계엽(KEK)

for the KOTO Collaboration 2017 KPS
Fall Meeting


J-PARC KOTO Experiment

- $\text{Br}(K_L \rightarrow \pi^0 \nu \bar{\nu}) = (2.8 \pm 0.4) \times 10^{-11}$ predicted by SM
- FCNC process in Standard model (Suppressed)
- Clean mode to explore the New Physics






CsI Calorimeter and Hermetic Veto Counters

New Pb/Scint Calorimeter




- Better suppression of background events



Inner Barrel



- 25 layers of 1-mm thick Pb sheet and 5-mm thick plastic scintillator
- Add $5X_0$ to $13.5X_0$




Cosmic Ray Test



- Attenuation curves fitted by two terms
 - P_2 describes the wavelength dependency
- Superior timing resolution of IB obtained by Cosmic-ray

$K_L \rightarrow \pi^0 \pi^0 \pi^0$ Reconstruction

Using 5 γ on Csl and 1 γ on Barrel




$$E_6 = \frac{M_\pi^2}{2E_5(1-\cos\theta)}$$




$$M_{K_L}^2 = \left(\sum_{i=1}^6 E_i\right)^2 - \left(\sum_{i=1}^6 \vec{p}_i\right)^2$$

- $K_L \rightarrow \pi^0 \pi^0 \pi^0$ decay samples with 5 γ s on Csl and 1 γ on Barrel
- Reconstruction of 2 π^0 from 4 γ s on Csl
- 1 γ Reconstruction from hit information of Barrel (timing and Module ID)
- Reconstruction of the third π^0 from 1 γ on Csl and 1 γ on Barrel

Reconstruction of Vertex X, Y




Iteration (10 times)






Red : M.C.
Black : Data(Run62)

Reconstruction Quality

Difference of Vertex Z







Difference of Incident Gamma Energy






$K_L \rightarrow \pi^0 \pi^0 \pi^0$ Monte Carlo Generation

Response Comparison




Vertex Time Difference

K_L Vertex Time






- Vertex Time Reconstruction with
 - Barrel
 - CsI Calorimeter
- Vertex Time Difference
 - Invariant

Timing Resolution from K_L signal



Sampling Fraction



- Sampling Calorimeter collects signal only from **Active**
 - Plastic scintillator
- **Passive** induces interaction with high Z number
 - Lead plate

Summary

- Reconstruction of $K_L \rightarrow \pi^0 \pi^0 \pi^0$ is done with 99.7% accuracy.
- Fine time and energy calibration are done with clear gamma selection.
- Timing Resolutions of Barrel Detector are evaluated using K_L signal.
 - p0 : 0.13[ns], p1 : 1.0[ns] for Inner Barrel
 - p0 : 0.03[ns], p1 : 2.1[ns] for Main Barrel
- Sampling Fractions of Barrel Detector are determined.
 - 0.29(IMB), 0.20(OMB), 0.31(IB)