

# A new charged particle detector for the KOTO experiment at J-PARC

**HongMin KIM for the KOTO Collaboration**

Division of Science Education, Jeonbuk National University, Jeonju 54896, Republic of Korea


E-mail: [recenter@naver.com](mailto:recenter@naver.com)

**Abstract.** We installed a new detector called Downstream Charged Veto(DCV) in order to suppress the  $K_L \rightarrow \pi^+\pi^-\pi^0$  decay background for the J-PARC KOTO experiment. Since the background is caused by non-detected charged pions passing through the beam hole of the electromagnetic calorimeter, the detector was installed in vacuum right behind. The DCV is composed of two plastic scintillator pipes read out by Multi Pixel Photon Counter(MPPC)s through wavelength shifting(WLS) fibers. From the test by using cosmic-rays during its fabrication, we obtained about 60 photoelectrons at the center of the DCV is about 60. After its installation, energy calibration was done with cosmic-rays surrounding the  $K_L$  beam, and its signal is identified by detectors surrounding the DCV.

## 1. Introduction

The KOTO experiment at J-PARC is searching for the  $K_L \rightarrow \pi^0\nu\bar{\nu}$  decay, which is one of the most sensitive probes to new physics beyond the standard model(SM). Its signature is a pair of photons from a  $\pi^0$  decay without any additional activity in a hermetic detector system surrounding the decay region. To detect this highly suppressed decay, expected at the  $3 \times 10^{-11}$  level, it is important to reject background events related to other kaon decay modes. At the single event sensitivity of  $1.30 \times 10^{-9}$  achieved by data collected in 2015, the number of  $K_L \rightarrow \pi^+\pi^-\pi^0$  background was estimated as  $0.05 \pm 0.02$  which corresponds to 2 at the SM sensitivity. The decay becomes background when charged pions passing through the beam hole are not detected due to their interaction with non-active materials. According to a Monte Carlo(MC) simulation, as shown in Fig. 1, there are three materials where the  $\pi^+$  and  $\pi^-$  interact and disappear. One is the membrane which separate decay region be evacuated as  $10^{-5}$  Pa from the detector region where is relatively low level of vacuum( $10^{-2}$  Pa). The other source is a pipe made of 0.5-mm-thick G-10, which prevents the membrane from drooping down to beam. The last one is a beam pipe made of 10-mm-thick aluminum for extending the highly evacuated decay region far from the calorimeter.


To reduce the  $K_L \rightarrow \pi^+\pi^-\pi^0$  background events, we have to detect the charged pions before they interact with non-active materials. For the purpose, we decided to install a new charged particle detector, DCV, inside a high vacuum region downstream of the electromagnetic calorimeter. To minimize the non-detected area, the DCV is placed as close as possible to the electromagnetic calorimeter.




**Figure 1.** Interacting point with  $\pi^+$  and  $\pi^-$ . Red dots indicate where the charged pions disappeared.

## 2. Structure of the DCV

The DCV consists of two successive square pipes, and each of them is made of 4 sheets of scintillators combined. The DCV1 is placed at 463 mm downstream from the calorimeter and inside the membrane. The DCV1 is a trapezoidal pipe, and the length of its upstream side is 157.5 mm and its downstream side is 171.5 mm. The beam direction length is 1410 mm. The DCV2 is placed at 76 mm downstream from the DCV1 and inside the aluminum beam pipe. The DCV2 is a square pipe, and the length of its upstream and downstream side is 173.5 mm. The beam direction length is 1460 mm. The scheme of the DCV is shown in Fig. 2. One module of the DCV consists of a 5-mm-thick plastic scintillator(EJ200, Eljen Technology) with 18 embedded WLS fibers(Y-11(200M), Kuraray). The diameter of WLS fiber is 1 mm. Due to very limited space for the DCV, we evaluated the new scheme of the light collection. In the new scheme of the light collection, the 4 MPPCs are directly attached to the surface of the scintillator through the light guide made of aluminum which is placed in the scintillator. The WLS fibers in the grooved scintillator are routed into the light guide. In this design, the WLS fibers are naturally bent to converge into the light guide. Figure 3 shows the light loss due to the curvature of the WLS fiber. We measured the light yield by MPPC of the LED light(430 nm) passing through the bent fiber. The light consistency was measured using PMT before injecting LED light into the fiber. The light loss increases rapidly if the radius is less than 20 mm.



**Figure 2.** Scheme of the DCV.



**Figure 3.** The light loss due to the curvature of the WLS fiber. The radius is a parameter of curvature.

49 **3. Fabrication Process**

50 *3.1. MPPC Gain Measurement and Fiber Test*


51 The 4 MPPCs belonging to each scintillator plate were applied the same operating voltage.  
 52 To group the MPPCs into four similar gain sets, we measured the MPPC (S13360-6050PE,  
 53 Hamamatsu)s single-photon gain using the LED light(430 nm). We also measured the light yield  
 54 of the WLS fibers. The LED light(430 nm) was injected on one side of the flagged WLS fiber  
 55 that we assigned before, and the MPPC was attached on the other side. After the measurement  
 56 of the light yield, we chose the WLS fiber with a high light yield.

57 *3.2. Making the scintillator pipe*


58 First, the WLS fibers were glued to the plastic scintillator using the optical cement(BC-600,  
 59 Saint-Gobain). The scintillators were dried for at least 48 hours. Second, the scintillators were  
 60 placed in a vacuum chamber. We extracted the outgas from the glued scintillators at less than  
 61 1 Pa, over 48 hours. Third, the scintillators were wrapped by a 12- $\mu$ m-thick aluminized film.  
 62 Next, the MPPCs were respectively attached on the light guide and fixed by aluminum plates.  
 63 After the cosmic-ray test, the scintillators were assembled to a square pipe.

64 *3.3. Cosmic-ray test*

65 To evaluate the light yield of the DCV, we measured the number of p.e. using cosmic-rays at  
 66 8 points, as shown in Fig. 4. At the center, the average number of p.e. for 1 MeV was 60.2  
 67 for the DCV1 and 58.6 for the DCV2. By fitting the data with an exponential function, the  
 68 attenuation length was found to be  $2469 \pm 165.1$  mm for the DCV1 and  $2566 \pm 166.0$  mm for  
 69 the DCV2.




**Figure 4.** The number of p.e. at each cosmic-ray trigger point



**Figure 5.** Calibration method for the DCV with 4 MPPCs. There are 3 types of normalization factors( $f_1$  : for each MPPC,  $f_2$  : for a pair of MPPC at upstream(downstream),  $f_4$  : for all MPPCs).

70 **4. Energy Calibration**

71 After the installation of the DCV in the KL beamline, we took the cosmic-ray data for an  
 72 energy calibration. The CC04 and CC05 surrounding the DCV were used as the trigger counter.  
 73 Figure 5 shows the diagram of the calibration method for the DCV with 4 MPPCs which were  
 74 shared one energy deposit. The energy response to the cosmic-ray of each module of the DCV  
 75 was used to derive the 3 types of normalization factors. Each normalization factor was obtained  
 76 from MIP peak by fitting the pulse height distribution. During the beam time from Feb. to  
 77 Apr. 2019, we collected the cosmic-ray data. Figure 6 shows how the calibration factor varies  
 78 during the period. The calibration factors tend to increase over time.



**Figure 6.** Calibration Factor over time for DCV1(a) and DCV2(b).

79 **5. Summary**

80 We fabricated and installed a new charged particle detector, the DCV, for further rejection of  
 81 the background events from the  $K_L \rightarrow \pi^+ \pi^- \pi^0$  decay. Based on the cosmic-ray test performed  
 82 during its fabrication, the light yield is about 60 p.e. for 1 MeV at the center of the DCV. We  
 83 established a method of its calibration by using cosmic-ray identified by detectors surrounding  
 84 the DCV. The energy calibration was done for the MeV scale. Studies on stability of its  
 85 performance during the beam time is undergoing.

86 **Acknowledge**

87 This work is supported by the National Research Foundation of Korea-2017R1A2B4006359,  
 88 and the JSPS KAKENHI Grant No. JP23224007.

89 **Reference**

90 [1] J.K. Ahn *et al.* (KOTO Collaboration) 2019, *Phys. Rev. Lett.* **122** 021802